



Valco Instruments Co. Inc.



Declaration of Conformity

Manufacturer: Valco Instruments Co. Inc.
P. O. Box 55603
7811 Westview Drive
Houston, Texas 77055 USA
Telephone: (713) 688-9345
Fax: (713) 688-3948
Email: valco@vici.com

Valco Instruments declares that the product specified herein

Type of equipment:	Gas calibration equipment
Model Number:	DynaCalibrator 150
Product description:	Calibration gas generator

is in compliance with the following:

EMC Standards: EN61326-1: 2006
CISPR 11 Conducted emissions Group 1, Class A December 9, 2005
CISPR 11 Radiated emissions Group 1, Class A December 7, 2005
EN61000-4-2:1995, A1:1998, A2 :2001
4kV/8kV December 19, 2005
EN61000-4-3:2002, A1:2002 3V/m December 12, 2005
EN61000-4-4:1995, A1:2001 .5kV/1kV December 7, 2005
EN61000-4-5:1995, A1:2001 .5kV/1kV December 7, 2005
EN61000-4-6:1996, A1:2001 3Vrms December 5, 2005
EN61000-4-11:2004 >95%-0.5p, 30%-25p, >95%-250p
December 6, 2005
EN61000-3-2:1995, A14:2000 Class A December 6, 2005
EN61000-3-3:2003 December 6, 2005

European Contact: VICI AG (Valco International)
Parkstrasse 2
CH-6214 Schenkon
Switzerland
Telephone: Int + 41-41-925-6200
Fax: Int + 41-41-925-6201
Email: info@vici.ch

S. D. Stearns, president

April 16, 2010

Date